



भारत सरकार Government of India
अंतरिक्ष विभाग Department of Space
विद्युत प्रकाशिकी तंत्र प्रयोगशाला
LABORATORY FOR ELECTRO-OPTICS SYSTEMS
प्रथम स्टेज, प्रथम क्रॉस /1ST STAGE, IST CROSS,
पीण्या औद्योगिक सम्पदा/ PEENYA INDUSTRIAL ESTATE,
बेंगलुरु /BENGALURU-560 058.
दूरभाष सं./ Phone No.080 28398836
फैक्स सं/ Fax No:080-28391964
ई मेल : Email:purchase@leos.gov.in

Ref :लियोस / LEOS/ पी टी /PT/07/2026-27/ दिनांक / Dated 22 July. 2026.

संशोधन संख्या : CORRIGENDUM NO.2 / Tender Posting on HQ Website -विज्ञापन संदर्भ सं /Advt. Ref. Number LEOS/PT/04/2026-27, दिनांक /Dated: 25 May 2026 and CORRIGENDUM NO.1 लियोस / LEOS/ पी टी /PT/05/2026-27/ दिनांक / Dated 11 June. 2026.

विषय / Subject: **LEOS 2026001018 01 for EOI FOR Supply, Installation, Demonstration and Training of Large Area Thin Film Coating System (LATCS)**

कृपया उपरोक्त निविदा को इसरो, ई-पोर्टल के दिनांक 25/05/2026 & 11/06/2026.

Kindly refer the subject e tender published in ISRO e portal dated 25.05.2026 & 11/06/2026.

अधोहस्ताक्षरी के द्वारा उक्त निविदा में निम्नलिखित परिवर्तन अधिसूचित किया जाता है।

The undersigned hereby notify the following changes in the subject e tender.

01	LEOS 2026001018 01	EOI FOR Supply, Installation, Demonstration and Training of Large Area Thin Film Coating System (LATCS)
क्रम सं / SI. No. (As per RFP)	बदले में / In Place of	यथा परिवर्तित / Read As
6.1	Inverted magnetron type Linear ion source shall be provided for pre-cleaning of the substrates Make: Genco	Inverted magnetron type Linear ion source shall be provided for pre-cleaning of the substrates Make: Genco or equivalent which meets or having better specifications shall be offered along with brand and make (OEM) details. Vendor shall include necessary datasheet and supporting documents.
7.1	Source Internal mounted linear magnetron sputtering source - Three numbers (03) Make: Genco	Source - Internal mounted linear/planar magnetron sputtering source - 3 Nos. Make: Genco or equivalent which meets or having better specifications shall be offered along with brand and make (OEM) details. Vendor shall include necessary datasheet and supporting documents.
7.5	Power source: Pulsed DC power supply unit, 10 kW pulsed Make: Advanced Energy	Power source: Pulsed DC power supply unit, 10 kW pulsed Make: Advanced Energy or equivalent which meets or having better specifications shall be offered along with brand and make (OEM) details. Vendor shall include necessary datasheet and supporting documents.
9.2	Operating range: 10E-6 to 2E-1 mbar without any differential pumping. Make: Genco, Model: OPTIX	Operating range: 10E-6 to 2E-1 mbar without any differential pumping. Make: Genco, Model: OPTIX or equivalent which meets or having better specifications shall be offered along with brand and make (OEM) details. Vendor shall include necessary datasheet and supporting documents.
28.7 (2)	Participation in the pre bid meeting is mandatory. Vendors who are not participating will not be considered for further evaluation.	Deleted.

मूल निविदा सूचना की अन्य सभी नियम और शर्तें अपरिवर्तित रहेगी।

All other terms and conditions of the original tender notice shall remain unchanged.

Sd/-
क्रय एवं भंडार अधिकारी
Purchase & Stores Officer
कृते एवं वास्ते भारत के राष्ट्रपति, क्रेता
For & on behalf of President of India,
The Purchaser



Laboratory for Electro-Optics Systems (LEOS), ISRO

Peenya Industrial Estate, Bengaluru-560 058

Pre-EOI meeting for Large Area Thin Film Coating System

Team LEOS invited all the participants for pre-EOI meeting and initiated the common queries session. Vendors from different organizations are expressed their clarification and shared their suggestions.

S.No	Vendor Queries	LEOS Feedback
1	LEOS to clarify if the mirror will have center hole	Depending on the telescope configuration some of the mirrors to be coated will have center hole. However, coating plant shall be designed to meet all types of optics. Coating uniformity to be met in entire 2000 mm diameter
2	LEOS to confirm if alternative magnetron sputter sources can be considered?	The magnetron source with meets EOI specifications shall be offered along with brand and make (OEM) details. Vendor shall include necessary datasheet and supporting documents.
3	LEOS to clarify if the sputtering targets are included in the supplier's scope of supply, or whether they will be arranged by LEOS?	Sputtering targets and spare targets shall be offered and it is included in the scope of supply.
4	LEOS to clarify the applicable payment terms for this project? In particular, please specify the expected payment schedule, including the payments associated with each project milestone	EOI refers to expression of interest for technical aspects. Commercial aspects will be part of tender document (RFP), which will be published during the next phase of the procurement process.
5	Vendor heritage aspects	The details will be presented in the RFP.
6	With respect to sub systems (like pumping system/ optical monitor/ etc.) options for more brands to be considered	The sub systems which meets or having better specifications shall be offered along with brand and make (OEM) details. Vendor shall include necessary datasheet and supporting documents.