

भारत सरकार/GOVERNMENT OF INDIA
अंतरिक्ष विभाग/DEPARTMENT OF SPACE
विक्रम साराभाई अंतरिक्ष केंद्र/VIKRAM SARABHAI SPACE CENTRE
तिरुवनंतपुरम/THIRUVANANTHAPURAM – 695 022

NO. VSSC/P/ADVT/278/2019

25.04.2019

दिनांक 16.04.2019 के विज्ञापन सं. वीएसएससी/पी/विद्यज्ञापन/278/2019 का शुद्धि-पत्र I
CORRIGENDUM I TO ADVT. NO. VSSC/P/ADVT/278/2019 DT. 16.04.2019

Changes in the compliance matrix of the following tender may please be noted :-

क्रम सं. Sl. No.	निविदा सं. Tender No.	वर्णन Description
04.	VSSC/PRSOPURCH/2019E1525901 [दो भाग / TWO PART]	अतिरिक्त एवं उप साधनों के साथ ग्लोव बक्स Glove Box with spares and accessories

CHANGES :

Sl No.	Sl No. in compliance matrix	Specification	Existing	Amended as
1.	1	Glove box: box dimension	1900 x 750 x 900 mm (min.)	1800 x 750 x 900 mm (min.)
2.	5	Vacuum pump: Pumping speed	20 m ³ /h	17 m ³ /h (min.)
3.	11	Oxygen Analyzer Make:	OEM/ GE/ TELEDYNE	Popular and reputed company make to be supplied
4.	12	Moisture Analyzer Make:	OEM/ GE/ TELEDYNE	Popular and reputed company make to be supplied

मूल निविदा सूचना की सभी अन्य निबंधन एवं शर्तें अपरिवर्तित रहेंगी।
All other terms and conditions of the original tender notice shall remain unchanged

हस्ताक्षरित/Sd/-

वरिष्ठ प्रधान, क्रय एवं भंडार / Sr. Head, Purchase & Stores